

Nisus NPWT Cleaning and Testing Log

Please refer to cleaning policy of NPWT device. Place initial under each section and sign signature portion at bottom with initials.

Date	Serial Number	Software Version	Equipment Cleaned	Suction Tested via Test Board	Power Cord Tested	Display Operating in Proper Manner	Alarm Log Cleared	Run Time Reset	Bag Sprayed with Bactericidal & Washed	Trouble Shooting Menu On	Battery at 100%	Pump Runs for 90 Seconds & Battery Above 80%	Alarms On	Service Sticker Placed	Once All Checks, Place in Patient Ready Section with Green Ready Sticker in Place
Comments:															